Test form for the External Quality Assessment (EQA) for laboratories participating in the European Antimicrobial Resistance Surveillance Network (EURGen-Net), 2026 – Enterobacterales

Participating laboratories can only submit results online via the webtool.
This form cannot be submitted.

Kindly note, due to the Material Transfer Agreements (MTAs) between SSI and the original providers of the isolate:

1. Strains received for the 2025 EURGen-Net EQA cannot be re-distributed further by the recipient laboratories.

2. It is not possible for DTU to distribute strains to laboratories after the EQA exercise, e.g. for confirmatory, training, or reference purposes.

It is recommended to store the strain in your strain collection (e.g. in a -80°C freezer), at least until you have reviewed your results from this EQA exercise. This will allow for repetition of species identification and AST, if needed, in light of your individual performance.

This form is designed to help participating laboratories prepare their results before submission in the EURGen-Net EQA webtool (https://EURL-PH-AMR.eqa.dtu.dk). The EURGen-Net phenotypic EQA is coordinated by DTU Food, in their capacity as a consortium member of the EURL-PH-AMR, in consultation with ECDC.

Enterobacterales- strain no.: _____

Enterobacterales- species: _____

The isolate should be considered as being obtained from a patient with a bloodstream infection.

Non-reported results will not be scored, but the antimicrobials and respective empty result sections will still be visible in the individual evaluation reports.

INFORMATION ABOUT THE METHOD

1. Which methodology did you use for species identification of this strain?

- ☐ Biochemical tests
- ☐ MALDI-TOF
- ☐ Whole-genome sequencing
- ☐ Other sequencing (e.g. 16S rRNA)
- ☐ PCR
- ☐ Other – Specify:_____

2. Which methodology did you mainly use for antimicrobial susceptibility testing (AST) of this strain?

- ☐ Automated system
- ☐ Disk/Tablet diffusion
- ☐ Broth microdilution
- ☐ Gradient test
- ☐ Macro broth dilution (tubes)
- ☐ Agar dilution
- ☐ Other – specify: _____

If you have used a different AST method for some antibiotics, please change the method for the relevant antimicrobial(s) below.

In the webtool, the type of method will by default be set to the information already provided above, and be aware that the settings below will change, if you later change method above.

Antimicrobial	Method
Amikacin	
Aztreonam	
Aztreonam-avibactam fixed 4 mg/L	
Cefiderocol	
Cefotaxime	
Ceftazidime	
Ceftazidime-avibactam fixed 4 mg/L	
Ceftolozane-tazobactam fixed 4 mg/L	
Ciprofloxacin	
Colistin	
Ertapenem	
Fosfomycin (*E. coli* only)	
Gentamicin	
Imipenem	
Imipenem-relebactam fixed 4 mg/L	
Levofloxacin	
Meropenem	

Antimicrobial	Method
Meropenem-vaborbactam fixed 8 mg/L	
Piperacillin-tazobactam fixed 4 mg/L	
Tigecycline (*E. coli* only)	
Tobramycin	
Trimethoprim-sulfamethoxazole (ratio 1:19)	

If you selected Automated system in the table above, please specify the instrument

- Microscan Walkaway
- Phoenix
- VITEK
- Other – specify: _____

If you selected Disk/Tablet diffusion in the table above, please specify the origin of the disks/tablets

- BD/BBL sensi disc
- Liofilchem
- MAST
- Neo sensitabs
- Oxoid
- Other– specify: _____

If you selected Disk/Tablet diffusion in the table above, please specify the origin of the agar

- BD BBL MH II Agar (Becton Dickinson)
- Biolife MH Agar II (Biolife Italiana)
- bioMerieux MHE Agar (bioMérieux)
- Bio-Rad MH Agar (Bio-Rad Laboratories)
- E&O Laboratories MH Agar (E&O Laboratories)
- Hardy Diagnostics MH Agar (Hardy Diagnostics)
- HiMedia MH Agar (HiMedia)
- HiMedia MH Agar no. 2 (HiMedia)
- Liofilchem MH II Agar (Liofilchem)
- Oxoid MH Agar (Thermo Scientific)
- Other – specify: _____

If you selected Broth microdilution in the table above, please specify the test and origin

- ComASP
- Liofilchem
- Sensititre MIC plates
- UMIC (Bruker)
- MIC plates prepared in-house
- Other– specify: _____

If you selected Broth microdilution in the table above, please specify the origin of the broth

- BD BBL
- Oxoid
- Sensititre
- Sigma-Aldrich
- Other – specify: _____

If you selected Gradient test in the table above, please specify the test and origin

- ○ E-test (bioMérieux)
- ○ MIC strip (Liofilchem)
- ○ Other – specify: _____

If you selected Gradient test in the table above, please specify the origin of the agar

- ○ BD BBL MH II Agar (Becton Dickinson)
- ○ Biolife MH Agar II (Biolife Italiana)
- ○ bioMerieux MHE Agar (bioMérieux)
- ○ Bio-Rad MH Agar (Bio-Rad Laboratories)
- ○ E&O Laboratories MH Agar (E&O Laboratories)
- ○ Hardy Diagnostics MH Agar (Hardy Diagnostics)
- ○ HiMedia MH Agar (HiMedia)
- ○ HiMedia MH Agar no. 2 (HiMedia)
- ○ Liofilchem MH II Agar (Liofilchem)
- ○ Oxoid MH Agar (Thermo Scientific)
- ○ Other – specify: _____

***3*. Which standard/guideline did you use when performing AST?**

☐ EUCAST – specify breakpoint table version: _____

☐ Other – specify: _____

RESULTS

Strain ID	Antimicrobial	Results and interpretation		
		≤ / = / >	MIC value (mg/L) or zone diameter (mm)	S / I / R / NA
	Amikacin			
	Aztreonam			
	Aztreonam-avibactam fixed 4 mg/L			
	Cefiderocol			
	Cefotaxime			
	Ceftazidime			
	Ceftazidime-avibactam fixed 4 mg/L			
	Ceftolozane-tazobactam fixed 4 mg/L			
	Ciprofloxacin			
	Colistin			
	Ertapenem			
	Fosfomycin (*E. coli* only)			
	Gentamicin			
	Imipenem			
	Imipenem-relebactam fixed 4 mg/L			

	Levofloxacin			
	Meropenem			
	Meropenem-vaborbactam fixed 8 mg/L			
	Piperacillin-tazobactam fixed 4 mg/L			
	Tigecycline (*E. coli* only)			
	Tobramycin			
	Trimethoprim-sulfamethoxazole (ratio 1:19)			